


# CLASSIC SERIES | ONYX™



BONE TYPES	All Bone Types Recommended for Hard Bone Type
PROSTHETICS PLATFORM	Internal hex
DESIGN FEATURES	<ul style="list-style-type: none"> <li>• Large surface area</li> <li>• Cylindrical thread and core body</li> <li>• Double threads with small step</li> <li>• Triple cutting flutes</li> </ul>
CLINICAL BENEFITS	<ul style="list-style-type: none"> <li>• Minimal pressure on hard bone</li> <li>• Maximum bone to implant contact area</li> <li>• Long term stability</li> <li>• Immediate loading – suitable for hard bone</li> </ul>

## ORDERING INFORMATION



Ø D (mm)	Ø D1 (mm)	L (mm)	Ref. No
3.3	2.4	8	NM-G3308
		10	NM-G3310
		11.5	NM-G3311
		13	NM-G3313
		16	NM-G3316
3.75	2.8	6	NM-G3706
		8	NM-G3708
		10	NM-G3710
		11.5	NM-G3711
		13	NM-G3713
		16	NM-G3716
4.2	3.2	6	NM-G4206
		8	NM-G4208
		10	NM-G4210
		11.5	NM-G4211
		13	NM-G4213
		16	NM-G4216
5.0	4.0	6	NM-G5006
		8	NM-G5008
		10	NM-G5010
		11.5	NM-G5011
		13	NM-G5013
		16	NM-G5016
6.0	5.0	6	NM-G6006
		8	NM-G6008
		10	NM-G6010
		11.5	NM-G6011
		13	NM-G6013


Cover Screw Included with all implants




NM-S5023

# CLASSIC SERIES | ONYX™

## RECOMMENDED STRAIGHT DRILL PROTOCOL



## RECOMMENDED STEP DRILL PROTOCOL



	Drill to mark osteotomy site		Drill osteotomy to implant		Drill osteotomy partially according to implant		Drill with countersink to prepare the crest
---	------------------------------	---	----------------------------	---	--	---	---

The recommended drill protocol procedure should not replace the dentist's/surgeon's judgment. The implants may be loaded for immediate function when good primary stability (above 35 Ncm) has been achieved and with appropriate occlusal loading.